

## Contents

1. Description
  - 1.1 Background information
  - 1.2 Applications
  - 1.3 Recommended reagent dilution
  - 1.4 Reagent requirements
2. Protocol
  - 2.1 (Optional) Lysis of whole blood
  - 2.2 General protocol for immunofluorescent staining
3. Example of immunofluorescent staining with CD19 CAR FMC63 Idiotype Antibody, REAfinity

## Warnings

Reagents contain sodium azide. Under acidic conditions sodium azide yields hydrazoic acid, which is extremely toxic. Azide compounds should be diluted with running water before discarding. These precautions are recommended to avoid deposits in plumbing where explosive conditions may develop.

## 1. Description

**This product is for research use only.**

**Components** 60 µL CD19 CAR FMC63 Idiotype Antibody, REAfinity

Product	Order no.
CD19 CAR FMC63 Idiotype Antibody, Biotin, REAfinity (clone: REA1297)	130-127-345
CD19 CAR FMC63 Idiotype Antibody, PE, REAfinity (clone: REA1297)	130-127-342
CD19 CAR FMC63 Idiotype Antibody, APC, REAfinity (clone: REA1297)	130-127-343
CD19 CAR FMC63 Idiotype Antibody, Vio® Bright B515, REAfinity (clone: REA1297)	130-127-344
CD19 CAR FMC63 Idiotype Antibody, Biotin, REAfinity (clone: REA1298)	130-127-349

**Capacity** For 30 tests or up to  $3 \times 10^7$  total cells.

**Product format** The reagent is supplied in buffer containing stabilizer and 0.05% sodium azide.

**Storage** Store protected from light at 2–8 °C. Do not freeze. The expiration date is indicated on the vial label.

## 1.1 Background information

The CD19 CAR FMC63 Idiotype Antibody, REAfinity has been developed for the detection of transduced T cells that are engineered to express chimeric antigen receptors (CARs) consisting of the scFV region of a CD19-specific monoclonal antibody, FMC63. The CD19 CAR FMC63 Idiotype Antibody, REAfinity (clone: REA1297 or REA1298) is an anti-idiotypic antibody. It recognizes the antigen recognition domain of FMC63 scFV from CD19 CAR T cells and contains a specifically mutated human IgG1 Fc region. The mutated human IgG1 Fc region of the CD19 CAR FMC63 Idiotype Antibody abolishes its binding to Fcγ receptors. This allows for background-free analysis and eliminates the need for additional blocking steps, such as using a FcR blocking reagent.

## 1.2 Applications

- Identification and enumeration of CD19 CAR<sup>+</sup> T cells by flow cytometry.

## 1.3 Recommended reagent dilution

The recommended dilution for CD19 CAR FMC63 Idiotype Antibody, REAfinity is **1:50 for up to  $10^6$  cells/100 µL**, e.g., 2 µL in a final staining volume of 100 µL for labeling of up to  $10^6$  cells and subsequent analysis by flow cytometry.

Cells should be stained prior to fixation, if formaldehyde is used as a fixative.

## 1.4 Reagent requirements

- PEB buffer: autoMACS® Running Buffer (# 130-091-221). Alternatively, prepare a solution containing phosphate-buffered saline (PBS), pH 7.2, 0.5% bovine serum albumin (BSA), and 2 mM EDTA by diluting MACS® BSA Stock Solution (# 130-091-376) 1:20 with autoMACS Rinsing Solution (# 130-091-222).
- (Optional) Red Blood Cell Lysis Solution (10×) (# 130-094-183) or (# 170-080-033, CE-IVD).
- (Optional) Double-distilled water (ddH<sub>2</sub>O).
- (Optional) 7-AAD Staining Solution (# 130-111-568 or # 170-080-032, CE-IVD).
- (Optional) Inside Fix, a component of the Inside Stain Kit (# 130-090-477).
- (Optional) If staining is done with a Biotin conjugate, additionally use Biotin Antibody, REAfinity (clone: REA746). It is recommended to use Biotin Antibody, PE, REAfinity (# 130-110-951). The total volume of the staining cocktail is 100 µL (including Biotin Antibody, REAfinity, 7-AAD Staining Solution, and additional fluorescently-conjugated antibodies). See table 1 for a recommendation of antibodies for the staining cocktail.

Product name	Clone
CD3 Antibody, anti-human, REAfinity	REA613
CD4 Antibody, anti-human, REAfinity	REA623
CD8 Antibody, anti-human, REAfinity	REA734
CD14 Antibody, anti-human, REAfinity	REA599
CD45 Antibody, anti-human, REAfinity	REA747

**Table 1:** Recommended antibodies to include in the staining panel.

## 2. Protocol

### 2.1 (Optional) Lysis of whole blood

- Dilute 10× Red Blood Cell Lysis Solution 1:10 with double-distilled water (ddH<sub>2</sub>O). For example, dilute 2 mL of 10× Red Blood Cell Lysis Solution with 18 mL of ddH<sub>2</sub>O.  
**▲ Note:** Do not dilute with deionized water. Store prepared 1× Red Blood Cell Lysis Solution at room temperature. Discard unused solution at the end of the day.
- Add an appropriate amount of whole blood to a suitable tube, e.g., add 1 mL of whole blood to a tube with 50 mL capacity.
- Add 1× Red Blood Cell Lysis Solution in 20-fold excess to the whole blood. For example, add 20 mL of 1× Red Blood Cell Lysis Solution to 1 mL of whole blood.
- Vortex immediately and thoroughly for 3 seconds and incubate for 10–20 minutes in the dark at room temperature (19–25 °C).
- Centrifuge at 300×g for 10 minutes. Remove supernatant.
- (Optional) Add more than 20× volume of PEB buffer of the initial cell sample volume, e.g., add 20 mL buffer when using 1 mL blood. Centrifuge at 300×g for 10 minutes. Remove supernatant.  
**▲ Note:** An additional washing step reduces staining background.
- Resuspend cell pellet in a suitable amount of buffer and proceed to immunofluorescent staining (chapter 2.2, step 4).

### 2.2 General protocol for immunofluorescent staining

**▲** Volumes given below are for up to 10<sup>6</sup> nucleated cells. When working with fewer than 10<sup>6</sup> cells, use the same volumes as indicated. When working with higher cell numbers, scale up all reagent volumes and total volumes accordingly (e.g. for 2×10<sup>6</sup> nucleated cells, use twice the volume of all indicated reagent volumes and total volumes).

#### Immunofluorescent staining with Biotin conjugates

**▲** Prepare a staining cocktail containing Biotin Antibody, REAfinity, 7-AAD Staining Solution for dead cell exclusion, and additional fluorochrome-conjugated antibodies. For examples refer to table 1. For details refer to the respective data sheets.

- Determine cell number.
- Centrifuge cell suspension at 300×g for 10 minutes. Aspirate supernatant completely.
- Resuspend up to 10<sup>6</sup> nucleated cells per 98 µL of buffer.
- Add 2 µL of the CD19 CAR FMC63 Idiotypic Antibody, REAfinity.
- Mix well and incubate for 10 minutes in the dark at room temperature (19–25 °C).

- Wash cells by adding 1 mL of buffer per 10<sup>6</sup> cells.
- Mix well and centrifuge at 300×g for 5 minutes at room temperature (19–25 °C). Aspirate supernatant completely.
- Add 100 µL of staining cocktail containing Biotin Antibody, REAfinity, 7-AAD Staining Solution, and additional fluorochrome-conjugated antibodies. Mix cells by pipetting up and down.
- Incubate for 10 minutes in the dark at room temperature (19–25 °C).
- Wash cells by adding 1 mL of buffer per 10<sup>6</sup> cells.
- Mix well and centrifuge at 300×g for 5 minutes at room temperature (19–25 °C). Aspirate supernatant completely.
- (Optional for fixation) Add 250 µL of buffer and 250 µL of Inside Fix to the cells and incubate for 20 minutes in the dark at room temperature (19–25 °C).
- (Optional for fixation) Add up to 2 mL of buffer.
- (Optional for fixation) Centrifuge cells at 300×g for 5 minutes at room temperature (19–25 °C). Aspirate supernatant completely.
- Resuspend cell pellet in a suitable amount of buffer for analysis by flow cytometry.

**▲ Note:** Acquire the samples within 1 hour after staining.

#### Immunofluorescent staining with PE, APC, or Vio Bright B515 conjugates


**▲** Prepare a staining cocktail containing CD19 CAR FMC63 Idiotypic Antibody, REAfinity, 7-AAD Staining Solution for dead cell exclusion, and additional fluorochrome-conjugated antibodies. For examples refer to table 1. For details refer to the respective data sheets.

- Determine cell number.
- Centrifuge cell suspension at 300×g for 10 minutes. Aspirate supernatant completely.
- Resuspend up to 10<sup>6</sup> nucleated cells per 50 µL of buffer.
- Add 50 µL staining cocktail containing 2 µL CD19 CAR FMC63 Idiotypic Antibody, 7-AAD Staining Solution, and additional fluorochrome-conjugated antibodies.
- Mix well by pipetting up and down and incubate for 10 minutes in the dark at room temperature (19–25 °C).
- Wash cells by adding 1 mL of buffer per 10<sup>6</sup> cells.
- Mix well and centrifuge at 300×g for 5 minutes at room temperature (19–25 °C). Aspirate supernatant completely.
- (Optional for fixation) Add 250 µL of buffer and 250 µL of Inside Fix to the cells and incubate for 20 minutes in the dark at room temperature (19–25 °C).
- (Optional for fixation) Add up to 2 mL of buffer.
- (Optional for fixation) Centrifuge cells at 300×g for 5 minutes at room temperature (19–25 °C). Aspirate supernatant completely.
- Resuspend cell pellet in a suitable amount of buffer for analysis by flow cytometry.

**▲ Note:** Acquire the samples within 1 hour after staining.

### 3. Example of immunofluorescent staining with CD19 CAR FMC63 Idiotypic Antibody, REAfinity

CD19 CAR<sup>+</sup> T cells, generated using the CliniMACS Prodigy<sup>®</sup> TCT process within 12 days, were stained with CD19 CAR FMC63 Idiotypic Antibody, REAfinity, Biotin Antibody-PE, CD3-FITC, and 7-AAD Staining Solution and analyzed by flow cytometry using the MACSQuant<sup>®</sup> Analyzer 10. Cell debris and dead cells were excluded from the analysis based on scatter signals and 7-AAD fluorescence. Shown are viable CD3<sup>+</sup> T cells.



Refer to [www.miltenyibiotec.com](http://www.miltenyibiotec.com) for all data sheets and protocols. Miltenyi Biotec provides technical support worldwide. Visit [www.miltenyibiotec.com](http://www.miltenyibiotec.com) for local Miltenyi Biotec Technical Support contact information.

### Legal notices

#### Limited product warranty

Miltenyi Biotec B.V. & Co. KG and/or its affiliate(s) warrant this product to be free from material defects in workmanship and materials and to conform substantially with Miltenyi Biotec's published specifications for the product at the time of order, under normal use and conditions in accordance with its applicable documentation, for a period beginning on the date of delivery of the product by Miltenyi Biotec or its authorized distributor and ending on the expiration date of the product's applicable shelf life stated on the product label, packaging or documentation (as applicable) or, in the absence thereof, ONE (1) YEAR from date of delivery ("Product Warranty"). Miltenyi Biotec's Product Warranty is provided subject to the warranty terms as set forth in Miltenyi Biotec's General Terms and Conditions for the Sale of Products and Services available on Miltenyi Biotec's website at [www.miltenyibiotec.com](http://www.miltenyibiotec.com), as in effect at the time of order ("Product Warranty"). Additional terms may apply. BY USE OF THIS PRODUCT, THE CUSTOMER AGREES TO BE BOUND BY THESE TERMS.

THE CUSTOMER IS SOLELY RESPONSIBLE FOR DETERMINING IF A PRODUCT IS SUITABLE FOR CUSTOMER'S PARTICULAR PURPOSE AND APPLICATION METHODS.

#### Technical information

The technical information, data, protocols, and other statements provided by Miltenyi Biotec in this document are based on information, tests, or experience which Miltenyi Biotec believes to be reliable, but the accuracy or completeness of such information is not guaranteed. Such technical information and data are intended for persons with knowledge and technical skills sufficient to assess and apply their own informed judgment to the information. Miltenyi Biotec shall not be liable for any technical or editorial errors or omissions contained herein.

All information and specifications are subject to change without prior notice. Please contact Miltenyi Biotec Technical Support or visit [www.miltenyibiotec.com](http://www.miltenyibiotec.com) for the most up-to-date information on Miltenyi Biotec products.

#### Licenses

This product and/or its use may be covered by one or more pending or issued patents and/or may have certain limitations. Certain uses may be excluded by separate terms and conditions. Please contact your local Miltenyi Biotec representative or visit Miltenyi Biotec's website at [www.miltenyibiotec.com](http://www.miltenyibiotec.com) for more information.

The purchase of this product conveys to the customer the non-transferable right to use the purchased amount of the product in research conducted by the customer (whether the customer is an academic or for-profit entity). This product may not be further sold. Additional terms and conditions (including the terms of a Limited Use Label License) may apply.

CUSTOMER'S USE OF THIS PRODUCT MAY REQUIRE ADDITIONAL LICENSES DEPENDING ON THE SPECIFIC APPLICATION. THE CUSTOMER IS SOLELY RESPONSIBLE FOR DETERMINING FOR ITSELF WHETHER IT HAS ALL APPROPRIATE LICENSES IN PLACE. Miltenyi Biotec provides no warranty that customer's use of this product does not and will not infringe intellectual property rights owned by a third party. BY USE OF THIS PRODUCT, THE CUSTOMER AGREES TO BE BOUND BY THESE TERMS.

#### Trademarks

autoMACS, CliniMACS Prodigy, MACS, MACSQuant, the Miltenyi Biotec logo, REAfinity, and Vio are registered trademarks or trademarks of Miltenyi Biotec and/or its affiliates in various countries worldwide. All other trademarks mentioned in this publication are the property of their respective owners and are used for identification purposes only.

Copyright © 2021 Miltenyi Biotec and/or its affiliates. All rights reserved.