

## Index

1. Description
  - 1.1 Principle of MACS® separation
  - 1.2 Background and product applications
  - 1.3 Reagent and instrument requirements
2. Protocol
  - 2.1 Sample preparation
  - 2.2 Magnetic labeling
  - 2.3 Magnetic separation
  - 2.4 (Optional) Evaluation of T cell purity
3. Example of a separation using the Pan T Cell Isolation Kit

## 1. Description

<b>Components</b>	<p><b>1 mL Pan T Cell Biotin-Antibody Cocktail:</b> Cocktail of biotin-conjugated monoclonal antibodies against CD11b, CD16, CD20, CD56, and CD66abce.</p> <p><b>2 mL Anti-Biotin MicroBeads:</b> MicroBeads conjugated to a monoclonal anti-biotin antibody (isotype: mouse IgG1).</p>
<b>Size</b>	For 10 <sup>9</sup> total cells, up to 100 separations
<b>Product format</b>	<p>The Biotin-Antibody Cocktail is supplied in a solution containing stabilizer and 0.05% sodium azide.</p> <p>The Anti-Biotin MicroBeads are supplied as a suspension containing stabilizer and 0.05% sodium azide.</p>
<b>Storage</b>	Store protected from light at 4–8 °C. Do not freeze. The expiration date is indicated on the vial label.

**This product is applicable for the separation of cells from rhesus monkey (*Macaca mulatta*). The antibodies have been tested to cross-react with cynomolgus monkey (*Macaca fascicularis*). Cross-reactivity with other non-human primates has not been tested.**

### 1.1 Principle of MACS® separation

Using the non-human primate Pan T Cell Isolation Kit, T cells are isolated by depletion of non-T cells (negative selection). Non-T cells are indirectly magnetically labeled with a cocktail of biotin-conjugated monoclonal antibodies, as primary labeling reagent, and anti-biotin monoclonal antibodies conjugated to MicroBeads, as secondary labeling reagent. The magnetically labeled non-T cells are depleted by retaining them on a MACS® Column in the magnetic field of a MACS Separator, while the unlabeled T cells pass through the column.

### 1.2 Background and product applications

The Pan T Cell Isolation Kit is an indirect magnetic labeling system for the isolation of untouched T cells from rhesus monkey peripheral blood mononuclear cells (PBMCs). Non-T cells, i.e. B cells, NK cells, monocytes and granulocytes, are indirectly magnetically labeled by using a cocktail of biotin-conjugated antibodies against CD11b, CD16, CD20, CD56, and CD66abce, and Anti-Biotin MicroBeads. Isolation of highly pure T cells is achieved by depletion of magnetically labeled cells.

#### Examples of applications

- Functional studies on T cells in which effects due to minimal antibody-cross-linking of cell surface proteins should be avoided.
- Studies on signal requirements for T cell activation, induction of T cell proliferation, induction of T cell energy, etc.
- Studies on signal transduction in T cells.
- Studies on regulation of T cell cytokine expression.

### 1.3 Reagent and instrument requirements

- Buffer (degassed): Prepare a solution containing PBS (phosphate buffered saline) pH 7.2, 0.5% BSA (bovine serum albumin) and 2 mM EDTA by diluting MACS BSA Stock Solution (# 130-091-376) 1:20 with autoMACS™ Rinsing Solution (# 130-091-222). Keep buffer cold (4–8 °C).
  - ▲ **Note:** EDTA can be replaced by other supplements such as anticoagulant citrate dextrose formula-A (ACD-A) or citrate phosphate dextrose (CPD). BSA can be replaced by other proteins such as human serum albumin, human serum or fetal calf serum. Buffers or media containing Ca<sup>2+</sup> or Mg<sup>2+</sup> are not recommended for use.
- MACS Columns and MACS Separators: Choose the appropriate MACS Separator and MACS Columns according to the number of labeled cells and to the number of total cells.

Column	max. number of labeled cells	max. number of total cells	Separator
MS	10 <sup>7</sup>	2×10 <sup>8</sup>	MiniMACS, OctoMACS, VarioMACS, SuperMACS
LS	10 <sup>8</sup>	2×10 <sup>9</sup>	MidiMACS, QuadroMACS, VarioMACS, SuperMACS
XS	10 <sup>9</sup>	2×10 <sup>10</sup>	SuperMACS
autoMACS	2×10 <sup>8</sup>	4×10 <sup>9</sup>	autoMACS

▲ **Note:** Column adapters are required to insert certain columns into VarioMACS™ Separator or SuperMACS™ Separator. For details, see MACS Separator data sheets.

- (Optional) Fluorochrome-conjugated antibodies (e.g. CD3-PE # 130-092-009; CD3-APC # 130-091-998; CD45-FITC # 130-091-898; Anti-Biotin-PE # 130-090-756; Anti-Biotin-APC # 130-090-856).
- (Optional) Pre-Separation Filters (# 130-041-407) to remove cell clumps.
- (Optional) PI (propidium iodide) or 7-AAD for the flow cytometric exclusion of dead cells.

## 2. Protocol

### 2.1 Sample preparation

When working with rhesus monkey anticoagulated peripheral blood or buffy coat, PBMCs should be isolated by density gradient centrifugation, as with human samples (see “General protocols” in the User Manuals or visit [www.miltenyibiotec.com](http://www.miltenyibiotec.com)).

▲ **Note:** Remove platelets after density gradient separation: resuspend cell pellet in buffer and centrifuge at 200×g for 10–15 minutes at 20 °C. Carefully remove supernatant. Repeat washing step and carefully remove supernatant.

When working with lymphoid tissues, prepare a single-cell suspension by a standard preparation method (see “General Protocols” in the User Manuals or visit [www.miltenyibiotec.com](http://www.miltenyibiotec.com)).

▲ **Note:** Dead cells may bind non-specifically to MACS MicroBeads. To remove dead cells, we recommend using density gradient centrifugation or the Dead Cell Removal Kit (# 130-090-101).



### 2.2 Magnetic labeling

▲ Work fast, keep cells cold, use pre-cooled solutions. This will prevent capping of antibodies on the cell surface and non-specific cell labeling.

▲ Volumes for magnetic labeling given below are for up to  $10^7$  total cells. When working with fewer than  $10^7$  cells, use same volumes as indicated. When working with higher cell numbers, scale up all reagent volumes and total volumes, accordingly (e.g. for  $2 \times 10^7$  total cells, use twice the volume of all indicated reagent volumes and total volumes).

▲ For optimal performance it is important to obtain a single cell suspension before magnetic separation. Pass cells through 30  $\mu\text{m}$  nylon mesh (Pre-Separation Filters # 130-041-407) to remove cell clumps which may clog the columns.

▲ Working on ice may require increased incubation times. Higher temperatures and/or longer incubation times lead to non-specific cell labeling.

1. Determine cell number.
2. Centrifuge cell suspension at 300×g for 10 minutes. Pipette off supernatant completely.
3. Resuspend cell pellet in 40  $\mu\text{L}$  of buffer per  $10^7$  total cells.
4. Add 10  $\mu\text{L}$  of Biotin-Antibody Cocktail per  $10^7$  total cells.
5. Mix well and incubate for 10 minutes at 4–8 °C.
6. Wash cells by adding 1–2 mL of buffer per  $10^7$  cells and centrifuge at 300×g for 10 minutes. Pipette off supernatant completely.
7. Add 80  $\mu\text{L}$  of buffer per  $10^7$  total cells.
8. Add 20  $\mu\text{L}$  of Anti-Biotin MicroBeads per  $10^7$  total cells.
9. Mix well and incubate for an additional 15 minutes at 4–8 °C.
10. Wash cells by adding 1–2 mL of buffer per  $10^7$  cells and centrifuge at 300×g for 10 minutes. Pipette off supernatant completely.
11. Resuspend up to  $10^8$  cells in 500  $\mu\text{L}$  of buffer.  
▲ **Note:** For higher cell numbers, scale up buffer volume accordingly.
12. Proceed to magnetic separation (2.3).



### 2.3 Magnetic separation

▲ Choose an appropriate MACS Column and MACS Separator according to the number of labeled cells and to the number of total cells (see table in section 1.3).

#### Magnetic separation with MS and LS Columns

1. Place column in the magnetic field of a suitable MACS Separator (see “Column data sheets”).
2. Prepare column by rinsing with appropriate amount of buffer:  
MS: 500  $\mu\text{L}$                       LS: 3 mL
3. Apply cell suspension onto the column.  
  
Allow the cells to pass through and collect effluent as fraction with unlabeled cells, representing the enriched T cell fraction.
4. Wash column with appropriate amount of buffer. Perform washing steps by adding buffer three times, each time once the column reservoir is empty.  
MS: 3×500  $\mu\text{L}$                       LS: 3×3 mL  
  
Collect entire effluent in the same tube as effluent of step 3. This fraction represents the enriched T cells.
5. (Optional) Elute retained cells outside of the magnetic field. This fraction represents the magnetically labeled non-T cells.

#### Magnetic separation with XS Columns

For instructions on the column assembly and the separation, refer to the “XS Column data sheet”.

#### Magnetic separation with the autoMACS™ Separator

▲ Refer to the “autoMACS User Manual” for instructions on how to use the autoMACS™ Separator.


1. Prepare and prime the autoMACS Separator.
2. Place tube containing the magnetically labeled cells in the autoMACS Separator. Choose program “Deplete”.
3. Collect the negative fraction (outlet port “neg1”). This fraction represents the enriched T cells.
4. (Optional) Collect positive fraction (outlet port “pos1”). This fraction represents the magnetically labeled non-T cells.

#### 2.4 (Optional) Evaluation of T cell purity

The purity of the enriched T cells can be evaluated by flow cytometry or fluorescence microscopy. Stain aliquots of the cell fractions with a fluorochrome-conjugated antibody against a T cell marker, e.g. CD3-PE (# 130-092-009) or CD3-APC (# 130-091-998), and CD45, e.g. CD45-FITC (# 130-091-898), as recommended in the respective data sheet. Analyze cells by flow cytometry or fluorescence microscopy. Labeling of non-T cells with the Biotin-Antibody Cocktail can be visualized by counter-staining with fluorochrome-conjugated anti-biotin antibodies, e.g. Anti-Biotin-PE (# 130-090-756) or Anti-Biotin-APC (# 130-090-856). Staining with fluorochrome-conjugated streptavidin is not recommended.

### 3. Example of a separation using the Pan T Cell Isolation Kit

Isolation of untouched T cells from rhesus monkey PBMCs using the Pan T Cell Isolation Kit and an LS Column. Cells are fluorescently stained with CD3-PE (# 130-092-009) and CD45-FITC (# 130-091-898). Plots show CD45<sup>+</sup> leukocytes. Cell debris and dead cells were excluded from the analysis based on scatter signals and PI fluorescence.



Refer to [www.miltenyibiotec.com](http://www.miltenyibiotec.com) for all data sheets and protocols. Miltenyi Biotec provides technical support worldwide. Visit [www.miltenyibiotec.com/local](http://www.miltenyibiotec.com/local) to find your nearest Miltenyi Biotec contact.

#### Warnings

Reagents contain sodium azide. Under acidic conditions sodium azide yields hydrazoic acid, which is extremely toxic. Azide compounds should be diluted with running water before discarding. These precautions are recommended to avoid deposits in plumbing where explosive conditions may develop.

### Legal notices

#### Limited product warranty

Miltenyi Biotec B.V. & Co. KG and/or its affiliate(s) warrant this product to be free from material defects in workmanship and materials and to conform substantially with Miltenyi Biotec's published specifications for the product at the time of order, under normal use and conditions in accordance with its applicable documentation, for a period beginning on the date of delivery of the product by Miltenyi Biotec or its authorized distributor and ending on the expiration date of the product's applicable shelf life stated on the product label, packaging or documentation (as applicable) or, in the absence thereof, ONE (1) YEAR from date of delivery ("Product Warranty"). Miltenyi Biotec's Product Warranty is provided subject to the warranty terms as set forth in Miltenyi Biotec's General Terms and Conditions for the Sale of Products and Services available on Miltenyi Biotec's website at [www.miltenyibiotec.com](http://www.miltenyibiotec.com), as in effect at the time of order ("Product Warranty"). Additional terms may apply. BY USE OF THIS PRODUCT, THE CUSTOMER AGREES TO BE BOUND BY THESE TERMS.

THE CUSTOMER IS SOLELY RESPONSIBLE FOR DETERMINING IF A PRODUCT IS SUITABLE FOR CUSTOMER'S PARTICULAR PURPOSE AND APPLICATION METHODS.

#### Technical information

The technical information, data, protocols, and other statements provided by Miltenyi Biotec in this document are based on information, tests, or experience which Miltenyi Biotec believes to be reliable, but the accuracy or completeness of such information is not guaranteed. Such technical information and data are intended for persons with knowledge and technical skills sufficient to assess and apply their own informed judgment to the information. Miltenyi Biotec shall not be liable for any technical or editorial errors or omissions contained herein.

All information and specifications are subject to change without prior notice. Please contact Miltenyi Biotec Technical Support or visit [www.miltenyibiotec.com](http://www.miltenyibiotec.com) for the most up-to-date information on Miltenyi Biotec products.

#### Licenses

This product and/or its use may be covered by one or more pending or issued patents and/or may have certain limitations. Certain uses may be excluded by separate terms and conditions. Please contact your local Miltenyi Biotec representative or visit Miltenyi Biotec's website at [www.miltenyibiotec.com](http://www.miltenyibiotec.com) for more information.

The purchase of this product conveys to the customer the non-transferable right to use the purchased amount of the product in research conducted by the customer (whether the customer is an academic or for-profit entity). This product may not be further sold. Additional terms and conditions (including the terms of a Limited Use Label License) may apply.

CUSTOMER'S USE OF THIS PRODUCT MAY REQUIRE ADDITIONAL LICENSES DEPENDING ON THE SPECIFIC APPLICATION. THE CUSTOMER IS SOLELY RESPONSIBLE FOR DETERMINING FOR ITSELF WHETHER IT HAS ALL APPROPRIATE LICENSES IN PLACE. Miltenyi Biotec provides no warranty that customer's use of this product does not and will not infringe intellectual property rights owned by a third party. BY USE OF THIS PRODUCT, THE CUSTOMER AGREES TO BE BOUND BY THESE TERMS.

#### Trademarks

autoMACS, MACS, MidiMACS, the Miltenyi Biotec logo, MiniMACS, OctoMACS, QuadroMACS, SuperMACS and VarioMACS are registered trademarks or trademarks of Miltenyi Biotec and/or its affiliates in various countries worldwide.

Copyright © 2020 Miltenyi Biotec and/or its affiliates. All rights reserved.