Miltenyi Biotec

MACS® GMP CD34-PE

Order no. 170-076-509

Contents

1. **Description**
2. **Applications**
3. **Warnings and precautions**
4. **Instructions for use**
5. **Glossary of symbols**

1. Description

Components	1 vial containing 5 mL MACS® GMP CD34-PE
Product format	MACS® GMP CD34-PE is supplied in phosphate-buffered saline (PBS), pH 7.4, and stabilizer containing 0.1% recombinant Human Serum Albumin (HSA).
Clone	AC101 (isotype: mouse IgG1)
Capacity	For labeling up to 1×10^9 total cells or lower cell numbers (e.g. 1×10^7) in the case of pre-enrichment with e.g. CliniMACS CD34 Reagent or CliniMACS CD133 Reagent.
Source	Mouse hybridoma cells
Endotoxin content	≤5 EU/mL as determined by kinetic Limulus Amebocyte Lysate (LAL) assay (Pharmacopoeia Europaea (Ph. Eur.)).
Stain Index	>2.5 as defined by $\frac{(\textit{mean positive cells}) - (\textit{mean negative cells})}{2 \times (\textit{standard deviation negative cells})}$
Sterility	Sterile. Manufactured aseptically, sterile filtered, filled aseptically.
Transport	At +2 °C to +8 °C (+36 °F to +46 °F).
Storage	Store MACS GMP CD34-PE protected from light at +2 °C to +8 °C (+36 °F to +46 °F) directly upon arrival.
Shelf life	The use-by-date is indicated on the vial label.

Disclaimer

MACS GMP Products are for *ex vivo* cell processing only, and are not intended for human *in vivo* applications. For regulatory status in the USA, please contact your local representative.

Quality statement

MACS GMP Products are manufactured and tested under a quality management system (ISO 13485) and are in compliance with relevant GMP guidelines. They are designed following the recommendations of USP <1043> on ancillary materials. The manufacturing and testing of this product is compliant with the requirements laid down in the Ph.Eur. Chapter 5.2.12.

The declaration of animal- or human-derived materials is given on the Certificate of Origin.

Background information

MACS GMP CD34-PE consists of a mouse monoclonal antibody against the CD34 antigen conjugated to phycoerythrin (PE). The CD34 antibody (clone AC101) recognizes the CD34 antigen, which is a single chain transmembrane glycoprotein expressed on human hematopoietic progenitor cells, endothelial progenitor cells, vascular endothelial cells, embryonic fibroblasts, and some cells in fetal and adult nervous tissue.

2. Applications

Fluorescent labeling of CD34 antigen on human cells.

3. Warnings and precautions

- The instructions for use must be followed.
- Do not inject or infuse the product directly into humans. Not for human application.
- When using this product, the national legislation and regulations must be followed. Any application of *ex vivo* processed target cells is exclusively within the responsibility of the user.
- For single use only. Do not reuse.
- Use undamaged and sealed vials only.
- Do not use after the use-by date indicated on the vial label.

33888/04 | Issued: 2022-11

Miltenyi Biotec B.V. & Co. KG
Friedrich-Ebert-Straße 68, 51429 Bergisch Gladbach, Germany
+49 2204 8306-0, +49 2204 85197
macs@miltenyibiotec.de
www.miltenyibiotec.com

Miltenyi Biotec Inc.
2303 Lindbergh Street, Auburn, CA 95602, USA
800 FOR MACS, +1 530 888 8871, +1 877 591 1060
macs@miltenyibiotec.com

4. Instructions for use

- For labeling cells and flow cytometric analysis the recommended antibody dilution for MACS GMP CD34-PE is 1:11. Up to 1×10^9 total cells or lower cell numbers (e.g. 1×10^7) in the case of pre-enrichment with e.g. CliniMACS CD34 (REF 171-01) or CliniMACS CD133 Reagent (REF 172-01) in 50 mL of CliniMACS PBS/EDTA Buffer are labeled with 1 vial of MACS GMP CD34-PE.
- For detailed instructions, refer to the Special Protocol for MACS® GMP labeling of cells and preparation for cell sort on MACSQuant® Tyto® on our website.

5. Glossary of symbols

	Manufacturer	+2 °C +8 °C	Temperature limit
REF	Order number		Keep away from sunlight.
P/N	Part number		Do not use if package is damaged.
LOT	Batch code		Phone
	Consult instruction for use		Fax
	Use-by date		E-mail
STERILE A	Sterilized using aseptic processing techniques		Website

This data sheet and corresponding information as well as special protocols can be found under **www.miltenyibiotec.com/170-076-509.**

Warranty

The products sold hereunder are warranted only to be free from defects in workmanship and material at the time of delivery to the customer. Miltenyi Biotec B.V. & Co. KG makes no warranty or representation, either expressed or implied, with respect to the fitness of a product for a particular purpose. There are no warranties, expressed or implied, which extend beyond the technical specifications of the products. Miltenyi Biotec B.V. & Co. KG's liability is limited to either replacement of the products or refund of the purchase price. Miltenyi Biotec B.V. & Co. KG is not liable for any property damage, personal injury or economic loss caused by the product.

CliniMACS, MACS, Milteny Biotec logo, MACSQuant, and Tyto are registered trademarks of Miltenyi Biotec B.V. & Co. KG and/or its affiliates in various countries worldwide.

Copyright © 2022 Miltenyi Biotec and/or its affiliates. All rights reserved.

33888/04 | Issued: 2022-11

Unless otherwise specifically indicated, Miltenyi Biotec products and services are for research use only and not for diagnostic or therapeutic use.