Miltenyi Biotec

StraightFrom® Buffy Coat CD8 MicroBead Kit
human

Order no. 130-114-978

Contents

1. Description
   - 1.1 Principle of the MACS® Separation
   - 1.2 Technical specifications of Whole Blood Columns
   - 1.3 Background information
   - 1.4 Applications
   - 1.5 Reagent and instrument requirements
2. Protocol
3. Example of a separation using the StraightFrom Buffy Coat CD8 MicroBeadKit, human

1. Description

This product is for research use only.

Components	**4 mL StraightFrom Buffy Coat CD8 MicroBeads, human:** MicroBeads conjugated to monoclonal anti-human CD8 antibodies (isotype: mouse IgG2a). **50 mL Whole Blood Column Elution Buffer** **12 Whole Blood Columns** columns and plungers, sterile packed.
Capacity	For one buffy coat (max. 80 mL) from max. 500 mL whole blood.
Product format	StraightFrom Buffy Coat CD8 MicroBeads, human are supplied in buffer containing stabilizer and 0.05% sodium azide. Whole Blood Column Elution Buffer contains stabilizer and 0.09% sodium azide.
Storage	Store StraightFrom Buffy Coat CD8 MicroBeads, human and Whole Blood Column Elution Buffer protected from light at +2 to +8 °C. Do not freeze. Store Whole Blood Columns dry at +10 to +35 °C and protected from light. The expiration date is indicated on the vial or box label.

1.1 Principle of the MACS Separation

First, the $CD8^+$ cells in a buffy coat sample are magnetically labeled with StraightFrom Buffy Coat CD8 MicroBeads, human. Then, the cell suspension is loaded onto a MACS Column, which is placed in the magnetic field of a MACS Separator. The magnetically labeled $CD8^+$ cells are retained within the column. The unlabeled cells run through; this cell fraction is thus depleted of $CD8^+$ cells. After removing the column from the magnetic field, the magnetically retained $CD8^+$ cells can be eluted as the positively selected cell fraction.

1.2 Technical specifications of Whole Blood Columns

- Columns are “flow stop” and do not run dry.
- Void volume: 600 µL. Reservoir volume: 7.5 mL.
- Whole Blood Columns are for single use only.
- Use Whole Blood Column Elution Buffer to elute positive cell fraction from column.

▲ To remove clumps and to prevent aggregates in the sample, pass cells through 30 µm nylon mesh (MACS SmartStrainers (30 µm), # 130-098-458) before separation.

1.3 Background information

The StraightFrom Buffy Coat CD8 MicroBead Kit, human has been developed for the positive selection of $CD8^+$ cells directly from one buffy coat by using the MultiMACS™ Cell24 Separator Plus or the QuadroMACS™ Separator. No sample preparation is required, including density gradient centrifugation or erythrocyte lysis. In human whole blood the CD8 antigen is expressed on cytotoxic T cells and on a subset of $CD16^+$ NK cells. The CD8 antigen acts as an accessory molecule in the recognition of MHC class I/peptide complexes by the TCR heterodimer on $CD8^+$ cytotoxic T cells.

1.4 Applications

- Isolation of $CD8^+$ cells from buffy coat. The purified $CD8^+$ cells are well suited for further flow cytometric, functional, or molecular analysis.

1.5 Reagent and instrument requirements

- Separation buffer: Prepare a solution containing phosphate-buffered saline (PBS), pH 7.2, 0.5% bovine serum albumin (BSA), and 2 mM EDTA by diluting MACS BSA Stock Solution (# 130-091-376) 1:20 with autoMACS® Rinsing Solution (# 130-091-222). Keep buffer cold (+2 to +8 °C). Alternatively, use autoMACS Running Buffer. Degas buffer before use, as air bubbles could block the column.

  ▲ **Note:** EDTA can be replaced by other supplements such as anticoagulant citrate dextrose formula-A (ACD-A) or citrate phosphate dextrose (CPD). BSA can be replaced by other proteins such as human serum albumin, human serum, or fetal bovine serum (FBS). Buffers or media containing Ca^{2+} or Mg^{2+} are not recommended for use.

Miltenyi Biotec B.V. & Co. KG
Friedrich-Ebert-Straße 68, 51429 Bergisch Gladbach, Germany
Phone +49 2204 8306-0, Fax +49 2204 85197
macsde@miltenyi.com
www.miltenyibiotec.com

- Table 1 provides an overview of MACS Columns and MACS Separators with links to the corresponding complete protocols. Use the link matching the MACS Column and MACS Separator used.

Table 1: MACS Columns and MACS Separators overview.

MACS Column (link to protocol)	Buffy coats	MACS Separator
Columns for manual separation		
Whole Blood Column	1 per 12 columns	QuadroMACS
Columns for instruments		
Whole Blood Column	1 per 12 columns	MultiMACS Cell24 Separator Plus

- Semi-automated separation:
  - MultiMACS Cell24 Separator Plus (# 130-098-637)
  - MultiMACS 12× Single-Column Adapter LS (# 130-108-816)
  - Single-well Deep Well Plates (# 130-114-966) or 24-well Deep Well Plates (# 130-110-500)
- Manual separation:
  - QuadroMACS Separator (# 130-090-976)
  - MACS Acrylic Tube Rack (# 130-041-406) or MACS 15 mL Tube Rack (# 130-091-052)
- (Optional) Fluorochrome-conjugated antibodies for flow cytometric analysis, e.g., CD3 Antibody, anti-human. For more information about antibodies refer to www.miltenyibiotec.com/antibodies.
- (Optional) Propidium Iodide Solution (# 130-093-233) or 7-AAD Staining Solution (# 130-111-568) for flow cytometric exclusion of dead cells.
- (Optional) MACS SmartStrainers (30 µm) (# 130-098-458) to remove cell clumps.
- (Optional) Red Blood Cell Lysis Solution (10×) (# 130-094-183)


2. Protocol

▲ The StraightFrom Buffy Coat MicroBead Kit, human has been developed for positive selection of target cells from buffy coat samples of up to 80 mL. When working with higher volumes, please contact our Technical Support team.

▲ Complete labeling and separation protocols are available online. Use the link provided in table 1, based on the used MACS Column and MACS Separator.

3. Example of a separation using the StraightFrom Buffy Coat CD8 MicroBead Kit, human

Separation of a buffy coat sample using the StraightFrom Buffy Coat CD8 MicroBeads, human and the MultiMACS Cell24 Separator Plus with the Single-Column Adapter and Whole Blood Columns. Cells were fluorescently stained with CD3-PE, CD8-APC, as well as CD45-VioBlue® and analyzed by flow cytometry using the MACSQuant® Analyzer. Cells were triggered via CD45-VioBlue cell debris and dead cells were excluded from the analysis based on scatter signals and propidium iodide fluorescence.



Refer to **www.miltenyibiotec.com** for all data sheets and protocols. Miltenyi Biotec provides technical support worldwide. Visit www.miltenyibiotec.com for local Miltenyi Biotec Technical Support contact information.

Legal notices

Limited product warranty

Miltenyi Biotec B.V. & Co. KG and/or its affiliate(s) warrant this product to be free from material defects in workmanship and materials and to conform substantially with Miltenyi Biotec's published specifications for the product at the time of order, under normal use and conditions in accordance with its applicable documentation, for a period beginning on the date of delivery of the product by Miltenyi Biotec or its authorized distributor and ending on the expiration date of the product's applicable shelf life stated on the product label, packaging or documentation (as applicable) or, in the absence thereof, ONE (1) YEAR from date of delivery ("Product Warranty"). Miltenyi Biotec's Product Warranty is provided subject to the warranty terms as set forth in Miltenyi Biotec's General Terms and Conditions for the Sale of Products and Services available on Miltenyi Biotec's website at www.miltenyibiotec.com, as in effect at the time of order ("Product Warranty"). Additional terms may apply. BY USE OF THIS PRODUCT, THE CUSTOMER AGREES TO BE BOUND BY THESE TERMS.

THE CUSTOMER IS SOLELY RESPONSIBLE FOR DETERMINING IF A PRODUCT IS SUITABLE FOR CUSTOMER'S PARTICULAR PURPOSE AND APPLICATION METHODS.

Unless otherwise specifically indicated, all Miltenyi Biotec products and services are for research use only and not for diagnostic or therapeutic use.

Technical information

The technical information, data, protocols, and other statements provided by Miltenyi Biotec in this document are based on information, tests, or experience which Miltenyi Biotec believes to be reliable, but the accuracy or completeness of such information is not guaranteed. Such technical information and data are intended for persons with knowledge and technical skills sufficient to assess and apply their own informed judgment to the information. Miltenyi Biotec shall not be liable for any technical or editorial errors or omissions contained herein.

All information and specifications are subject to change without prior notice. Please contact Miltenyi Biotec Technical Support or visit www.miltenyibiotec.com for the most up-to-date information on Miltenyi Biotec products.

Licenses

This product and/or its use may be covered by one or more pending or issued patents and/or may have certain limitations. Certain uses may be excluded by separate terms and conditions. Please contact your local Miltenyi Biotec representative or visit Miltenyi Biotec's website at www.miltenyibiotec.com for more information.

The purchase of this product conveys to the customer the non-transferable right to use the purchased amount of the product in research conducted by the customer (whether the customer is an academic or for-profit entity). This product may not be further sold. Additional terms and conditions (including the terms of a Limited Use Label License) may apply.

CUSTOMER'S USE OF THIS PRODUCT MAY REQUIRE ADDITIONAL LICENSES DEPENDING ON THE SPECIFIC APPLICATION. THE CUSTOMER IS SOLELY RESPONSIBLE FOR DETERMINING FOR ITSELF WHETHER IT HAS ALL APPROPRIATE LICENSES IN PLACE. Miltenyi Biotec provides no warranty that customer's use of this product does not and will not infringe intellectual property rights owned by a third party. BY USE OF THIS PRODUCT, THE CUSTOMER AGREES TO BE BOUND BY THESE TERMS.

Trademarks

MACS, MACSQuant, the Miltenyi Biotec logo, MultiMACS, QuadroMACS, StraightFrom, and VioBlue are registered trademarks or trademarks of Miltenyi Biotec B.V. & Co. KG and/or its affiliates in various countries worldwide. All other trademarks mentioned in this publication are the property of their respective owners and are used for identification purposes only.

Copyright © 2026 Miltenyi Biotec and/or its affiliates. All rights reserved.

Unless otherwise specifically indicated, all Miltenyi Biotec products and services are for research use only and not for diagnostic or therapeutic use.

140-005-183.04