

## Contents

1. Description
  - 1.1 Principle of the Nuclei Extraction Buffer
  - 1.2 Background information
  - 1.3 Applications
  - 1.4 Reagent and instrument requirements
2. Protocol
  - 2.1 Sample, equipment, and reagent preparation
  - 2.2 Nuclei extraction protocol
  - 2.3 Labeling for microscopic or flow cytometric analysis of nuclei samples

## 1. Description

This product is for research use only.

<b>Components</b>	100 mL Nuclei Extraction Buffer
<b>Size</b>	For 25 extractions.
<b>Storage</b>	Store protected from light at 2–8 °C. Do not freeze. The expiration date is indicated on the label.

### 1.1 Principle of the Nuclei Extraction Buffer

Tissue samples from various sources and species can be dissociated into single nuclei suspensions by combining mechanical tissue dissociation and cell lysis. The tissue sample is added to a gentleMACS™ C Tube together with the Nuclei Extraction Buffer and RNase inhibitor. The sample is dissociated using the gentleMACS Dissociator.

After dissociation, the sample is applied to a filter to remove any remaining larger particles from the single-nuclei suspension. Nuclei are collected by centrifugation and the nuclei pellet is resuspended and filtered again.

### 1.2 Background information

The Nuclei Extraction Buffer and the corresponding gentleMACS Program have been optimized for a gentle, rapid, and effective generation of single-nuclei suspensions from fresh and frozen tissue samples. This reliable and standardized protocol enables the recovery of high yield single-nuclei suspensions that can be used for a wide variety of studies including proteomic analysis, FACS analysis, and single-nuclei gene expression analysis.

### 1.3 Applications

- Dissociation of fresh, frozen (including flash-frozen and OCT embedded) tissue samples from various tissues and species into single-nuclei suspensions for subsequent molecular analyses.
- The provided protocol has been validated for a wide range of tissues, including brain, liver, heart, lung, pancreas, spleen and kidney, as well as tumor tissues including human melanoma, breast, pancreatic, colon and prostate cancer and grafted mouse tumors.

### 1.4 Reagent and instrument requirements

- Centrifuge
- MACS® SmartStrainers (70 µm) (# 130-098-462)
- MACS SmartStrainers (30 µm) (# 130-098-458)
- gentleMACS Octo Dissociator (# 130-095-937) or gentleMACS Octo Dissociator with Heaters (# 130-096-427)
- gentleMACS C Tubes (# 130-093-237, # 130-096-334)
- RNase inhibitor (e.g. murine RNase Inhibitor from New England Biolabs)
- (Optional) Eppendorf® LoBind® Tubes 5 mL

### Additional requirements for working with mouse brain

- MACS SmartStrainers (100 µm) (# 130-098-463)

### Additional requirements for microscopic or flow cytometric analysis of single-nuclei suspensions (refer to protocol 2.3)

- Nuclei staining dye, e.g., 7-AAD Staining Solution (# 130-111-568), DAPI Staining Solution (# 130-111-570), or DRAQ5 Staining Solution (# 130-117-343).
- (Optional) Cell sorter, e.g., MACSQuant® Tyto® Sorter (# 130-103-931)

## 2. Protocol

▲ For details on the use of the gentleMACS Dissociators, refer to the gentleMACS Dissociators user manuals.

▲ Please make sure that the gentleMACS Octo Dissociator runs with the latest software update. The **4C\_nuclei\_1** program is included in the Miltenyi program folder from software version GM\_V02.H21 (gentleMACS Octo Dissociator with Heaters) or GM\_V02.R11 (gentleMACS Octo Dissociator) onwards. The latest software update can be requested at <https://www.miltenyibiotec.com/lp/2018/gentlemacs-software-update.html>.

▲ Pipette gently and slowly during all nuclei resuspension steps to minimize alteration of the extracted nuclei.

## 2.1 Sample, equipment, and reagent preparation

▲ Pre-cool centrifuge, buffers, and consumables with sample contact (e.g. gentleMACS C Tube and SmartStrainers) at 4 °C.

▲ **Note:** It is recommended to pre-cool the C Tube containing prepared lysis buffer overnight at 4 °C. Ideally, perform all nuclei extraction steps in a cold room.

### Preparation of lysis buffer

Per extraction add RNase inhibitor (final concentration 0.2 U/μL) to pre-cooled 4 mL Nuclei Extraction Buffer.

### Preparation of resuspension buffer

Prepare an appropriate amount of resuspension buffer, e.g., nuclei separation buffer for further enrichment of nuclei using Anti-Nucleus MicroBeads (# 130-132-997). For more information refer to the respective data sheet.

Per extraction the following volumes of resuspension buffer are recommended:

Amount/type of tissue	Resuspension buffer
up to 50 mg	1.5 mL
50–100 mg	3 mL
100–200 mg	6 mL
Mouse brain hemisphere	8 mL

**Table 1:** Overview of resuspension buffer volumes.

▲ **Note:** It is strongly recommended to adjust the nuclei concentration to  $<5 \times 10^6$ /mL to avoid nuclei aggregation.

▲ **Note:** When working with tissues with high intrinsic RNase activity (e.g. pancreas), the RNase inhibitor concentration can be increased in both prepared lysis and resuspension buffer in order to prevent RNA degradation.

▲ **Note:** If only DNA-based downstream applications are to be performed, RNase inhibitor can be left out.

▲ **Note:** Some genomic applications require the use of a different resuspension buffer to ensure optimal performance. Follow the recommendations of the genomic application provider.

### Preparation of tissue sample

Resect tissue and cut into pieces of up to 200 mg. To store tissue, freeze them on dry ice or use liquid nitrogen.

▲ **Note:** It is recommended to cut tissue before freezing. If frozen tissue pieces are  $>200$  mg, cut them without letting samples thaw.

▲ **Note:** In minor cases, e.g., kidney, it is recommended to cut up the tissue into pieces of  $<50$  mg to ensure complete dissociation.

▲ **Note:** When working with OCT-embedded tissue, remove excess OCT without letting samples thaw.

## 2.2 Nuclei extraction protocol

▲ Optimal extraction performance depends on starting material (tissue type, amount of starting material) and downstream application and might require protocol optimization.

▲ Work fast until tissue is dissociated and keep samples on ice at all steps.

▲ Volumes given below are for up to 200 mg tissue per extraction.

1. Add 2 mL ice-cold lysis buffer to each pre-cooled gentleMACS C Tube.

2. Transfer tissue pieces to the gentleMACS C Tube containing lysis buffer and directly proceed the following steps until samples are dissociated.

▲ **Note:** Do not let frozen sample thaw before dissociation as endogenous RNase might degrade RNA.

3. Close gentleMACS C Tube and place it on the gentleMACS Dissociator.

▲ **Note:** Close C Tube tightly beyond the first resistance.

4. Run gentleMACS Program **4C\_nuclei\_1** on the gentleMACS Dissociator.

5. After termination of the program, detach C Tube from the gentleMACS Dissociator and place the C Tube immediately on ice.

▲ **Note:** Depending on tissue type, additional 5–10 minutes incubation on ice after dissociation might be needed to optimize cell lysis. Over-lysis should be avoided as it could lead to nuclei aggregation and nuclei damage.

6. Apply nuclei suspension to a MACS SmartStrainer (70 μm) placed on a 15 mL tube.

▲ **Note:** For brain samples, use a MACS SmartStrainer (100 μm).

▲ **Note:** To increase nuclei recovery, the use of low binding tubes is recommended at this and all following steps.

7. Wash MACS SmartStrainer with 2 mL ice-cold lysis buffer.

▲ **Note:** Alternatively, for maximum cell recovery, rinse used C Tube with ice-cold lysis buffer before adding to the MACS SmartStrainer.

8. Discard MACS SmartStrainer and centrifuge nuclei suspension at  $300 \times g$  at 4 °C for 5 minutes. Carefully aspirate supernatant completely.

9. Resuspend nuclei pellet with ice-cold resuspension buffer by slowly and gently pipetting the sample up and down for 10 times.

10. Apply nuclei suspension to a MACS SmartStrainer (30 μm) placed on a 15 mL tube.

▲ **Note:** When working with myelin-rich brain, a myelin removal step is strongly recommended to remove myelin debris.

11. Collect nuclei suspension and proceed immediately with downstream applications, e.g., single nuclei RNA seq or nuclei sorting.

▲ **Note:** In case debris amount is still too high for downstream application, purify the nuclei suspension by using a cell sorter, e.g. MACSQuant Tyto.

### 2.3 Labeling for microscopic or flow cytometric analysis of nuclei samples

1. Add 7-AAD Staining Solution (final concentration of 0.525 µg/mL), DAPI Staining Solution (final concentration of 0.25 µg/mL), or DRAQ5 (10 µM) Staining Solution to a small fraction of the nuclei suspension. Mix by inverting gently.
2. Incubate 5 minutes at 4 °C.
3. Load sample on a hemocytometer or on a flow cytometer, e.g., MACSQuant Analyzer.
4. Analyze sample according to manufacturer's recommendation.

Refer to [www.miltenyibiotec.com](http://www.miltenyibiotec.com) for all data sheets and protocols. Miltenyi Biotec provides technical support worldwide. Visit [www.miltenyibiotec.com](http://www.miltenyibiotec.com) for local Miltenyi Biotec Technical Support contact information.

## Legal notices

### Limited product warranty

Miltenyi Biotec B.V. & Co. KG and/or its affiliate(s) warrant this product to be free from material defects in workmanship and materials and to conform substantially with Miltenyi Biotec's published specifications for the product at the time of order, under normal use and conditions in accordance with its applicable documentation, for a period beginning on the date of delivery of the product by Miltenyi Biotec or its authorized distributor and ending on the expiration date of the product's applicable shelf life stated on the product label, packaging or documentation (as applicable) or, in the absence thereof, ONE (1) YEAR from date of delivery ("Product Warranty"). Miltenyi Biotec's Product Warranty is provided subject to the warranty terms as set forth in Miltenyi Biotec's General Terms and Conditions for the Sale of Products and Services available on Miltenyi Biotec's website at [www.miltenyibiotec.com](http://www.miltenyibiotec.com), as in effect at the time of order ("Product Warranty"). Additional terms may apply. BY USE OF THIS PRODUCT, THE CUSTOMER AGREES TO BE BOUND BY THESE TERMS. THE CUSTOMER IS SOLELY RESPONSIBLE FOR DETERMINING IF A PRODUCT IS SUITABLE FOR CUSTOMER'S PARTICULAR PURPOSE AND APPLICATION METHODS.

### Technical information

The technical information, data, protocols, and other statements provided by Miltenyi Biotec in this document are based on information, tests, or experience which Miltenyi Biotec believes to be reliable, but the accuracy or completeness of such information is not guaranteed. Such technical information and data are intended for persons with knowledge and technical skills sufficient to assess and apply their own informed judgment to the information. Miltenyi Biotec shall not be liable for any technical or editorial errors or omissions contained herein.

All information and specifications are subject to change without prior notice. Please contact Miltenyi Biotec Technical Support or visit [www.miltenyibiotec.com](http://www.miltenyibiotec.com) for the most up-to-date information on Miltenyi Biotec products.

### Licenses

This product and/or its use may be covered by one or more pending or issued patents and/or may have certain limitations. Certain uses may be excluded by separate terms and conditions. Please contact your local Miltenyi Biotec representative or visit Miltenyi Biotec's website at [www.miltenyibiotec.com](http://www.miltenyibiotec.com) for more information.

The purchase of this product conveys to the customer the non-transferable right to use the purchased amount of the product in research conducted by the customer (whether the customer is an academic or for-profit entity). This product may not be further sold. Additional terms and conditions (including the terms of a Limited Use Label License) may apply.

CUSTOMER'S USE OF THIS PRODUCT MAY REQUIRE ADDITIONAL LICENSES DEPENDING ON THE SPECIFIC APPLICATION. THE CUSTOMER IS SOLELY RESPONSIBLE FOR DETERMINING FOR ITSELF WHETHER IT HAS ALL APPROPRIATE LICENSES IN PLACE. Miltenyi Biotec provides no warranty that customer's use of this product does not and will not infringe intellectual property rights owned by a third party. BY USE OF THIS PRODUCT, THE CUSTOMER AGREES TO BE BOUND BY THESE TERMS.

### Trademarks

autoMACS, gentleMACS, MACS, MACSQuant, the Miltenyi Biotec logo, and NeuroBrew are registered trademarks or trademarks of Miltenyi Biotec and/or its affiliates in various countries worldwide. All other trademarks mentioned in this publication are the property of their respective owners and are used for identification purposes only.

Copyright © 2023 Miltenyi Biotec and/or its affiliates. All rights reserved.