Miltenyi Biotec

MACS® GMP CD25-PE-Biotin

Order no. 170-076-505

Contents

1. **Description**
2. **Applications**
3. **Warnings and precautions**
4. **Instructions for use**
5. **Glossary of symbols**

1. Description

Components	1 vial containing 5 mL MACS® GMP CD25-PE-Biotin
Product format	MACS GMP CD25-PE-Biotin is supplied in phosphate-buffered saline (PBS), pH 7.4, and stabilizer containing 0.1 % recombinant human serum albumin (rHSA).
Clone	3G10 (isotype: mouse IgG1)
Capacity	For labeling up to 1×10^{10} total cells.
Source	Mouse hybridoma cells
Endotoxin content	≤5 EU/mL as determined by kinetic Limulus Amebocyte Lysate (LAL) assay (Pharmacopoeia Europaea (Ph. Eur.)).
Stain Index	>2.5 as defined by $\frac{(\textit{mean positive cells}) - (\textit{mean negative cells})}{2 \times (\textit{standard deviation negative cells})}$
Sterility	Sterile. Manufactured aseptically, sterile filtered, filled aseptically.
Transport	At +2 °C to +8 °C (+36 °F to +46 °F)
Storage	Store MACS GMP CD25-PE-Biotin protected from light at +2 °C to +8 °C (+36 °F to +46 °F) directly upon arrival.
Shelf life	The use-by-date is indicated on the vial label.

Disclaimer

MACS GMP Products are for *ex vivo* cell processing only, and are not intended for human *in vivo* applications. For regulatory status in the USA, please contact your local representative.

Quality statement

MACS GMP Products are manufactured and tested under a quality management system (ISO 13485) and are in compliance with relevant GMP guidelines. They are designed following the recommendations of USP <1043> on ancillary materials. The manufacturing and testing of this product is compliant with the requirements laid down in the Ph.Eur. Chapter 5.2.12.

The declaration of animal- or human-derived materials is given on the Certificate of Origin.

Background information

MACS GMP CD25-PE-Biotin consists of a mouse monoclonal antibody against the human CD25 antigen coupled to a tandem conjugate of phycoerythrin (PE) and Biotin. The CD25 antibody (clone 3G10) recognizes a 55 kDa glycoprotein also known as the low-affinity interleukin-2 receptor alpha chain (IL-2Rα). This molecule reveals three epitope regions: A, B and C. Clone 3G10 recognizes epitope region A.

CD25 is expressed on activated T and B cells, on macrophages, and on a subset of non-activated $CD4^+$ regulatory T cells.

2. Applications

Fluorescent labeling of CD25 antigen on human cells.

3. Warnings and precautions

- The instructions for use must be followed.
- Do not inject or infuse the product directly into humans. Not for human application.
- When using this product, the national legislation and regulations must be followed. Any application of *ex vivo* processed target cells is exclusively within the responsibility of the user.
- For single use only. Do not reuse.
- Use undamaged and sealed vials only.
- Do not use after the use-by date indicated on the vial label.

33862/06 | Issued: 2022-11


Miltenyi Biotec B.V. & Co. KG
Friedrich-Ebert-Straße 68, 51429 Bergisch Gladbach, Germany
+49 2204 8306-0, +49 2204 85197
macs@miltenyibiotec.de
www.miltenyibiotec.com

Miltenyi Biotec Inc.
2303 Lindbergh Street, Auburn, CA 95602, USA
800 FOR MACS, +1 530 888 8871, +1 877 591 1060
macs@miltenyibiotec.com

4. Instructions for use

- For labeling cells and flow cytometric analysis as well as magnetic enrichment the recommended antibody dilution for MACS GMP CD25-PE-Biotin is 1:20. For optimal staining performance vortexing of the MACS GMP CD25-PE-Biotin is required prior to use. Up to 1×10^{10} total cells in 95 mL of CliniMACS PBS/EDTA Buffer are labeled with 1 vial of MACS GMP CD25-PE-Biotin.
- For magnetic enrichment of CD25 positive cells using the CliniMACS System, refer to the Special Protocol Magnetic enrichment of CD25 positive cells using MACS GMP CD25-PE-Biotin and to the Special Protocol for MACS® GMP labeling of cells and preparation for cell sort on MACSQuant® Tyto® on our website.

5. Glossary of symbols

	Manufacturer		Temperature limit
REF	Order number		Keep away from sunlight.
P/N	Part number		Do not use if package is damaged.
LOT	Batch code		Phone
	Consult instruction for use		Fax
	Use-by date		E-mail
STERILE A	Sterilized using aseptic processing techniques		Website

This data sheet and corresponding information as well as special protocols can be found under **www.miltenyibiotec.com/170-076-505.**

Warranty

The products sold hereunder are warranted only to be free from defects in workmanship and material at the time of delivery to the customer. Miltenyi Biotec B.V. & Co. KG makes no warranty or representation, either expressed or implied, with respect to the fitness of a product for a particular purpose. There are no warranties, expressed or implied, which extend beyond the technical specifications of the products. Miltenyi Biotec B.V. & Co. KG's liability is limited to either replacement of the products or refund of the purchase price. Miltenyi Biotec B.V. & Co. KG is not liable for any property damage, personal injury or economic loss caused by the product.

CliniMACS, MACS, the Miltenyi Biotec Logo, MACSQuant, and Tyto are registered trademarks of Miltenyi Biotec B.V. & Co. KG and/or its affiliates in various countries worldwide.

Copyright © 2022 Miltenyi Biotec and/or its affiliates. All rights reserved.

33862/06 | Issued: 2022-11

Unless otherwise specifically indicated, Miltenyi Biotec products and services are for research use only and not for diagnostic or therapeutic use.