Miltenyi Biotec

MACS® GMP Cell Differentiation Bag

Cell Differentiation Bag – 100	REF 170-076-400
Cell Differentiation Bag – 250	REF 170-076-401
Cell Differentiation Bag – 500	REF 170-076-402
Cell Differentiation Bag – 1000	REF 170-076-404
Cell Differentiation Bag – 3000	REF 170-076-405

Contents

1. **Description**
2. **Applications**
3. **Warnings and precautions**
4. **Instructions for use**
5. **Glossary of symbols**

1. Description

Components 1 unit contains 5 Cell Differentiation Bags with volumes of 100 mL, 250 mL, 500 mL, 1000 mL, or 3000 mL (please see appropriate order number and table below).

For labeling purpose, sample identification tags and connection strips are provided.

Size	Product	Volume/Area	Minimal fill volume
1 unit	5× Cell Differentiation Bag – 100	100 mL/116 cm^2	20 mL
1 unit	5× Cell Differentiation Bag – 250	250 mL/156 cm^2	30 mL
1 unit	5× Cell Differentiation Bag – 500	500 mL/230 cm^2	50 mL
1 unit	5× Cell Differentiation Bag – 1000	1000 mL/337 cm^2	120 mL
1 unit	5× Cell Differentiation Bag – 3000	3000 mL/672 cm^2	220 mL

Sterility Sterile and non-pyrogenic.

Transport Room temperature

Storage Store at room temperature.

Shelf life Do not use after use-by date as given on the package.

Disclaimer

MACS GMP Products are for *ex vivo* cell processing only, and are not intended for human *in vivo* applications. For regulatory status in the USA, please contact your local representative.

Quality statement

MACS GMP Products are manufactured and tested under a quality management system (ISO 13485) and are in compliance with relevant GMP guidelines. They are designed following the recommendations of USP <1043> on ancillary materials.

2. Applications

- The transparent MACS GMP Cell Differentiation Bags are intended for the *in vitro* cultivation and expansion of cells from heterogenous hematologic cell populations, for example, for the generation of dendritic cells from CD14-enriched monocytes.

  The transparent cell culture bag allows for the cell culture process to be monitored by microscopy.

3. Warnings and precautions

- The instructions for use must be followed.
- When using this product, the national legislation and regulations must be followed. Any application of *ex vivo* processed target cells is exclusively within the responsibility of the user.
- For single use only. Do not reuse.
- Use undamaged and sealed product only.
- Do not use after the use-by date indicated on the product label.
- Cell Differentiation Bags should not be centrifuged.
- Do not use for cryopreservation.
- Writing on or labeling the Cell Differentiation Bag surface is not recommended.
- Surface disinfection: The bags are compatible with aldehyde and peroxide-based disinfection reagents. However, alcohol and quaternary ammonium complexes may compromise the integrity.
- Organic solvents, e.g., dimethyl sulfoxide (DMSO), may cause damage to Cell Differentiation Bags.

36701/09 | Issued: 2022-11

Miltenyi Biotec B.V. & Co. KG
Friedrich-Ebert-Straße 68, 51429 Bergisch Gladbach, Germany
+49 2204 8306-0, +49 2204 85197
macs@miltenyibiotec.de
www.miltenyibiotec.com

Miltenyi Biotec Inc.
2303 Lindbergh Street, Auburn, CA 95602, USA
800 FOR MACS, +1 530 888 8871, +1 877 591 1060
macs@miltenyibiotec.com

4. Instructions for use

- Use aseptic techniques.
- Be sure to carefully remove air from the bag when filling.
- Careful disinfection of the septum before and after usage is recommended.
- The transfer or removal of fluids can be performed by using the female luer lock connection at the end of the tubing (fig. 1) or by using the tubing for sterile docking (tubing PVC. inner/outer diameter 2.5/4.1 mm).
- Observe the instructions for use supplied by the manufacturer of the tube welding device.
- The septum can be used to apply or remove fluids using a syringe and needle (e.g. cytokines or sample for analysis).




Figure 1: Cell Differentiation Bag.

5. Glossary of symbols

	Manufacturer	STERILE R	Sterilized using irradiation
REF	Order number		Do not re-use
P/N	Part number		Phone
LOT	Batch code		Fax
	Consult instruction for use		E-mail
	Use-by date		Website
	Do not use if package is damaged.		

This data sheet and corresponding information can be found under **www.miltenyibiotec.com/170-076-400**.

Warranty

The products sold hereunder are warranted only to be free from defects in workmanship and material at the time of delivery to the customer. Miltenyi Biotec B.V. & Co. KG makes no warranty or representation, either expressed or implied, with respect to the fitness of a product for a particular purpose. There are no warranties, expressed or implied, which extend beyond the technical specifications of the products. Miltenyi Biotec B.V. & Co. KG's liability is limited to either replacement of the products or refund of the purchase price. Miltenyi Biotec B.V. & Co. KG is not liable for any property damage, personal injury or economic loss caused by the product.

MACS and the Miltenyi Biotec logo are registered trademarks or trademarks of Miltenyi Biotec B.V. & Co. KG and/or its affiliates in various countries worldwide.

Copyright © 2022 Miltenyi Biotec and/or its affiliates. All rights reserved.

36701/09 | Issued: 2022-11

Unless otherwise specifically indicated, Miltenyi Biotec products and services are for research use only and not for diagnostic or therapeutic use.