

25 M Tubes  
100 M Tubes

130-093-236  
130-096-335

## Contents

### 1. Description

- 1.1 Background information
- 1.2 Technical specifications
- 1.3 Reagent and instrument requirements
- 1.4 Applications

### 1. Description

This product is for research use only.

**Components** 25 M Tubes, sterile and single packed  
or  
100 M Tubes, sterile packed as 4×25 tubes.

**Storage** Store tubes dry and protected from light at room temperature (19–25 °C).  
The expiration date is indicated on the box label.  
Do not use after this date.

#### 1.1 Background information

The orange-colored gentleMACS M Tubes are designed for automated homogenization using the gentleMACS Dissociators. M Tubes are used for very efficient homogenization of tissues and cells to isolate subcellular material, such as RNA, DNA, and proteins or pathogens. The homogenization procedure is carried out in a closed system, thus avoiding cross-contamination and facilitating sterile handling.

The cap of the M Tube bears a rotor and a stator. An appropriate buffer is pipetted into the opened M Tube and the sample material is added. The tube is tightly closed with the cap. The tube is attached upside down to the sleeves of the gentleMACS Dissociators. It has to be ensured that the sample material is located in the area of the rotor/stator. The homogenization process is started by running an appropriate gentleMACS Program. After processing, the M Tube is detached from the gentleMACS Dissociators and the homogenized sample is removed from the M Tube.

#### 1.2 Technical specifications

- Sample sizes
  - 500 µL – maximal 10 mL total volume
  - 20–4000 mg of tissue

▲ **Note:** The tubes are suitable for a broad range of tissues, for example, spleen, liver, or brain. However, very hard material, such as bone, may damage the tubes.
- Dimensions
  - Length: 87 mm
  - Diameter: 30 mm
- Recommended working temperature is between 4 °C and 40 °C.

- Maximal centrifugal acceleration  
Relative centrifugal force (RCF) = 3000×g
- The tubes are for single use only.
- To close the M Tubes, turn the cap clockwise beyond the first resistance.

#### 1.3 Reagent and instrument requirements

- gentleMACS Dissociator (# 130-093-235), gentleMACS Octo Dissociator (# 130-095-937), or gentleMACS Octo Dissociator with Heaters (# 130-096-427)
- (Optional) µMACS™ mRNA Starting Kit (# 130-075-202), µMACS One-step cDNA Starting Kit (# 130-091-989), µMACS One-step cDNA Labeling Starting Kit (# 130-092-521), or µMACS One-step T7 Template Starting Kit (# 130-092-943).

#### 1.4 Applications

M Tubes are designed for the homogenization of tissues or cells to isolate subcellular material, such as RNA, DNA, and proteins or pathogens.

For further details and more applications, refer to [www.miltenyibiotec.com](http://www.miltenyibiotec.com). The gentleMACS Protocols are also online available at [www.miltenyibiotec.com/protocols](http://www.miltenyibiotec.com/protocols).

Refer to [www.miltenyibiotec.com](http://www.miltenyibiotec.com) for all data sheets and protocols. Miltenyi Biotec provides technical support worldwide. Visit [www.miltenyibiotec.com](http://www.miltenyibiotec.com) for local Miltenyi Biotec Technical Support contact information.

## Legal notices

### Limited product warranty

Miltenyi Biotec B.V. & Co. KG and/or its affiliate(s) warrant this product to be free from material defects in workmanship and materials and to conform substantially with Miltenyi Biotec's published specifications for the product at the time of order, under normal use and conditions in accordance with its applicable documentation, for a period beginning on the date of delivery of the product by Miltenyi Biotec or its authorized distributor and ending on the expiration date of the product's applicable shelf life stated on the product label, packaging or documentation (as applicable) or, in the absence thereof, ONE (1) YEAR from date of delivery ("Product Warranty"). Miltenyi Biotec's Product Warranty is provided subject to the warranty terms as set forth in Miltenyi Biotec's General Terms and Conditions for the Sale of Products and Services available on Miltenyi Biotec's website at [www.miltenyibiotec.com](http://www.miltenyibiotec.com), as in effect at the time of order ("Product Warranty"). Additional terms may apply. BY USE OF THIS PRODUCT, THE CUSTOMER AGREES TO BE BOUND BY THESE TERMS. THE CUSTOMER IS SOLELY RESPONSIBLE FOR DETERMINING IF A PRODUCT IS SUITABLE FOR CUSTOMER'S PARTICULAR PURPOSE AND APPLICATION METHODS.

### Technical information

The technical information, data, protocols, and other statements provided by Miltenyi Biotec in this document are based on information, tests, or experience which Miltenyi Biotec believes to be reliable, but the accuracy or completeness of such information is not guaranteed. Such technical information and data are intended for persons with knowledge and technical skills sufficient to assess and apply their own informed judgment to the information. Miltenyi Biotec shall not be liable for any technical or editorial errors or omissions contained herein.

All information and specifications are subject to change without prior notice. Please contact Miltenyi Biotec Technical Support or visit [www.miltenyibiotec.com](http://www.miltenyibiotec.com) for the most up-to-date information on Miltenyi Biotec products.

### Licenses

This product and/or its use may be covered by one or more pending or issued patents and/or may have certain limitations. Certain uses may be excluded by separate terms and conditions. Please contact your local Miltenyi Biotec representative or visit Miltenyi Biotec's website at [www.miltenyibiotec.com](http://www.miltenyibiotec.com) for more information.

The purchase of this product conveys to the customer the non-transferable right to use the purchased amount of the product in research conducted by the customer (whether the customer is an academic or for-profit entity). This product may not be further sold. Additional terms and conditions (including the terms of a Limited Use Label License) may apply.

CUSTOMER'S USE OF THIS PRODUCT MAY REQUIRE ADDITIONAL LICENSES DEPENDING ON THE SPECIFIC APPLICATION. THE CUSTOMER IS SOLELY RESPONSIBLE FOR DETERMINING FOR ITSELF WHETHER IT HAS ALL APPROPRIATE LICENSES IN PLACE. Miltenyi Biotec provides no warranty that customer's use of this product does not and will not infringe intellectual property rights owned by a third party. BY USE OF THIS PRODUCT, THE CUSTOMER AGREES TO BE BOUND BY THESE TERMS.

### Trademarks

gentleMACS, the Miltenyi Biotec logo, and  $\mu$ MACS are registered trademarks or trademarks of Miltenyi Biotec and/or its affiliates in various countries worldwide.

Copyright © 2023 Miltenyi Biotec and/or its affiliates. All rights reserved.