Miltenyi Biotec

REAlease® CD25 MicroBead Kit
human

Order no. 130-133-522

Contents

1. Description
   - 1.1 Principle of the REAlease MACS® Separation
   - 1.2 Background information
   - 1.3 Applications
   - 1.4 Reagent and instrument requirements
2. Protocol
   - 2.1 Protocol overview
   - 2.2 Sample preparation
   - 2.3 Magnetic labeling
   - 2.4 Magnetic separation and removal of magnetic labeling
   - 2.5 Magnetic labeling and separation using autoMACS® Separators
   - 2.6 (Optional) Removal of the REAlease Complex and second magnetic labeling with REAlease MicroBeads
   - 2.7 (Optional) Second magnetic labeling with MACS MicroBeads
3. Example of a separation using the REAlease CD25 MicroBead Kit

1. Description

This product is for research use only.

Components	**1 mL REAlease CD25-Biotin, human** **5 mL REAlease Anti-Biotin MicroBeads (CD25, human)** **4 mL REAlease Bead Release Reagent (50×)** **4 mL REAlease Release Reagent** **4 mL REAlease Stop Reagent**
Capacity	For 10^9 total cells, up to 100 separations.
Product format	REAlease Stop Reagent is supplied in buffer containing 0.05% sodium azide. All other reagents are supplied in buffer containing stabilizer and 0.05% sodium azide.
Storage	Store protected from light at 2–8 °C. Do not freeze. The expiration date is indicated on the vial label.

1.1 Principle of the REAlease MACS Separation

The REAlease Technology relies on recombinantly engineered antibody fragments to label specific cell surface markers. The fragments are developed to have low affinity for epitopes. However, when fragments are multimerized as a REAlease Biotin Complex (i.e., REAlease CD25-Biotin, human) they bind to epitopes with high avidity. REAlease Technology can control the multimer/monomer state of antibody fragments. With this technology a controlled label release is possible where monomerized antibody fragments dissociate from the cell surface, enabling users to obtain bead- and label-free cells.
First, the target cells in a peripheral blood mononuclear cell (PBMC) population are labeled with REAlease CD25-Biotin (REAlease Biotin Complex). Subsequently, REAlease Anti-Biotin MicroBeads (CD25, human) bind to the REAlease Biotin Complex. Then, the cell suspension is loaded onto a MACS Column, which is placed in the magnetic field of a MACS Separator. The magnetically labeled cells are retained within the column. The unlabeled non-target cells flow through; this cell fraction is thus depleted of $CD25^+$ cells. To increase purity, a second column run is needed. Therefore, cells are eluted directly from the first onto the second, equilibrated column. After washing the column again and removing the column from the magnetic field, the target cells are eluted using the REAlease Bead Release Reagent, which simultaneously removes the MicroBeads from the cells. Finally, during the subsequent incubation with the REAlease Release Reagent, the REAlease Biotin Complex monomerizes and dissociates from the cell surface leaving the cells free of all labels.

1.2 Background information

The REAlease CD25 MicroBead Kit has been developed for the isolation of $CD25^+$ cells from PBMCs.
The REAlease CD25 MicroBead Kit is an indirect magnetic labeling system that allows to obtain cells free of MicroBeads and REAlease Biotin Complex.
CD25, the low-affinity interleukin-2 receptor alpha chain (IL-2Rα), is expressed on T regulatory cells as well as on activated T and B cells and monocytes.
For isolation of $CD25^+$ cell subsets, the REAlease CD25 MicroBead Kit can be combined with REAlease CD4 MicroBead Kit.

1.3 Applications

- Isolation of $CD25^+$ label free cells.
- Isolation of label-free regulatory T cells

Miltenyi Biotec B.V. & Co. KG
Friedrich-Ebert-Straße 68, 51429 Bergisch Gladbach, Germany
Phone +49 2204 8306-0, Fax +49 2204 85197
macsde@miltenyi.com
www.miltenyibiotec.com

1.4 Reagent and instrument requirements

- Separation buffer: Prepare a solution containing phosphate-buffered saline (PBS), pH 7.2, 0.5% bovine serum albumin (BSA), and 2 mM EDTA by diluting MACS BSA Stock Solution (# 130-091-376) 1:20 with autoMACS Rinsing Solution (# 130-091-222). Use buffer at room temperature (19–25 °C). Store buffer cold (2–8 °C). Degas buffer before use, as air bubbles could block the column.

  ▲ **Note:** BSA can be replaced by other proteins such as human serum albumin, human serum, or fetal bovine serum (FBS). Buffers or media containing Ca^{2+} or Mg^{2+} are not recommended for use.

- REAlease Bead Release buffer: Prepare a 1:50 dilution of REAlease Bead Release Reagent (50×), e.g., for 1 mL add 20 µL of REAlease Bead Release Reagent to 980 µL of separation buffer.

  ▲ **Note:** Use freshly prepared buffer the same day. Store at room temperature.

  ▲ **Note:** Prepare 1 mL per MS Column and 5 mL per LS Column.

- MACS Columns and MACS Separators: $CD25^+$ cells can be enriched by using MS or LS Columns. Depletion can also be performed by using the MultiMACS™ Cell24 Separator Plus or autoMACS Columns on the autoMACS NEO Separator.

Column	Max. number of labeled cells	Max. number of total cells	Separator
Positive selection			
MS	10^7	2×10^8	MiniMACS, OctoMACS, SuperMACS II
LS	10^8	2×10^9	MidiMACS, QuadroMACS, SuperMACS II
autoMACS	2×10^8	4×10^9	autoMACS NEO Separator
Multi-24 Column Block (per column)	10^8	10^9	MultiMACS Cell24 Separator Plus


  ▲ **Note:** Column adapters are required to insert certain columns into SuperMACS™ II Separators. For details refer to the respective MACS Separator data sheet.

  ▲ **Note:** If separating with LS Columns and the MultiMACS Cell24 Separator Plus use the Single-Column Adapter. Refer to the user manual for details.

- (Optional) Fluorochrome-conjugated CD25 antibodies for flow cytometric analysis, e.g., CD25 Antibody, anti-human, APC. For more information about antibodies refer to www.miltenyibiotec.com/antibodies.
- (Optional) Propidium Iodide Solution (# 130-093-233), 7-AAD Staining Solution (# 130-111-568), or Viobility™ Fixable Dyes (# 130-109-812, # 130-109-814, # 130-109-816) for flow cytometric exclusion of dead cells.
- (Optional) Pre-Separation Filters (30 µm) (# 130-041-407) to remove cell clumps.

2. Protocol

2.1 Protocol overview



2.2 Sample preparation

When working with anticoagulated peripheral blood or buffy coat, peripheral blood mononuclear cells (PBMCs) should be isolated by using a MACS PBMC Isolation Kit or by density gradient centrifugation, for example, using Ficoll-Paque™.

▲ **Note:** To remove platelets after density gradient separation, resuspend cell pellet in buffer and centrifuge at 200×g for 10–15 minutes at 20 °C. Carefully aspirate supernatant. Repeat washing step.

For details refer to the data sheet or the protocols section at www.miltenyibiotec.com/protocols.



2.3 Magnetic labeling

▲ The recommended incubation temperature is at room temperature (19–25 °C).

▲ Volumes for magnetic labeling given below are for up to 10^7 total cells. When working with fewer than 10^7 cells, use the same volumes as indicated. When working with higher cell numbers, scale up all reagent volumes and total volumes accordingly (e.g. for 2×10^7 total cells, use twice the volume of all indicated reagent volumes and total volumes).

▲ For optimal performance it is important to obtain a single-cell suspension before magnetic labeling. Pass cells through 30 µm nylon mesh (Pre-Separation Filters (30 µm), # 130-041-407) to remove cell clumps which may clog the column. Moisten filter with buffer before use.

1. Determine cell number.
2. Centrifuge cell suspension at 300×g for 10 minutes. Aspirate supernatant completely.
3. Resuspend cell pellet in 40 µL of separation buffer per 10^7 total cells.
4. Add 10 µL of REAlease CD25-Biotin per 10^7 total cells.
5. Mix well and incubate for 5 minutes.
6. Add 50 µL of REAlease Anti-Biotin MicroBeads (CD25, human) per 10^7 total cells.
7. Mix well and incubate for 5 minutes.
8. (Optional) Add staining antibodies, e.g., C25 Antibody, anti-human, APC, and incubate according to manufacturer's recommendation.

   ▲ **Note:** These staining antibodies cannot be removed from the cells.
9. Dilute up to 5×10^7 cells in a total volume of 500 µL with separation buffer.

   ▲ **Note:** For volumes larger than 500 µL a dilution is not needed.
10. Proceed to magnetic separation (2.4).



2.4 Magnetic separation and removal of magnetic labeling

▲ Choose an appropriate MACS Column and MACS Separator according to the number of total cells and the number of $CD25^+$ cells. For details refer to the table in section 1.4.

▲ Always wait until the column reservoir is empty before proceeding to the next step.

▲ The recommended incubation temperature is at room temperature (19–25 °C).

Magnetic separation with two MS or two LS Columns

1. Prepare two columns by rinsing with the appropriate amount of separation buffer:

   MS: 500 µL LS: 3 mL
2. Place the first column in the magnetic field of a suitable MACS Separator. For details refer to the respective MACS Column data sheet.
3. Apply cell suspension onto the first column. Collect flow-through containing unlabeled cells.
4. Wash the first column with the appropriate amount of separation buffer. Collect unlabeled cells that pass through and combine with the flow-through from step 3.

   MS: 3×500 µL LS: 3×3 mL

   ▲ **Note:** Perform washing steps by adding buffer aliquots as soon as the column reservoir is empty.
5. Remove the first column from the separator.
6. Place the second column in the MACS Separator and put the first column on top of the second column.
7. Add the appropriate amount of separation buffer onto the first column. Immediately flush out target cells directly onto the second column by firmly pushing the plunger into the first column.

   MS: 1 mL LS: 5 mL
8. Remove the first column from the second column.
9. Wash second column with the appropriate amount of separation buffer.

   MS: 3×500 µL LS: 3×3 mL

   ▲ **Note:** Perform washing steps by adding buffer aliquots as soon as the column reservoir is empty.
10. Remove the second column from the separator and place it on a suitable collection tube.
11. (Optional) If removal of MicroBeads is not required, add the appropriate amount of separation buffer onto the second column. Immediately flush out target cells by firmly pushing the plunger into the column. Eluted cells are ready for downstream applications, e.g., flow cytometric analysis.

    MS: 1 mL LS: 5 mL
12. For removal of MicroBeads proceed with step 13.

13. Add the appropriate amount of prepared REAlease Bead Release buffer (refer to chapter 1.4) onto the second column. Immediately flush out target cells by firmly pushing the plunger into the column.

    MS: 1 mL LS: 5 mL

14. Mix well and incubate for 10 minutes.
15. Cells are now free from MicroBeads and ready for analysis and downstream applications.
16. (Optional) Proceed either to
    - 2.6 Removal of REAlease Complex and second magnetic labeling with REAlease MicroBeads

      or proceed to
    - 2.7 Second magnetic labeling with MACS MicroBeads.

Magnetic separation with the MultiMACS Cell24 Separator

Refer to the the MultiMACS Cell Separator user manual for instructions on how to use the MultiMACS Cell24 Separator.

2.5 Magnetic labeling and separation using autoMACS Separators

▲ Refer to the user manual and the short instructions for instructions on how to use the autoMACS Separators.

▲ Buffers used for operating the autoMACS Separators should have a temperature of ≥10 °C.

▲ Place tubes in the following Chill Rack positions:

position A = sample, position B = unlabeled (negative) fraction,

position C = labeled (positive) fraction.

2.5.1 Magnetic labeling and separation using the autoMACS NEO Separator

▲ The autoMACS NEO Separator enables stage loading to extend column capacity for selected reagents, minimizing the need to divide larger samples.

▲ The autoMACS NEO Separator provides three programs for the REAlease CD25 MicroBead Kit: purity, recovery, and high purity. The purity and recovery programs run on one column and are compatible for $CD25^+$ cell separation. The high purity program runs on two columns and it is intended for regulatory T cell separation.

▲ For more information on selecting alternative separation programs, stage loading–compatible reagents, autolabeling-compatible reagents, and the minimal and maximal volumes for each reagent and Chill Rack, refer to www.miltenyibiotec.com/automacs-neo-sample-processing.

Fully automated magnetic labeling and separation

1. Prepare and prime the instrument.
2. Place the Chill Rack and MACS Reagent Rack 8 on the MACS MiniSampler S.
3. Select the same Chill Rack and MACS Reagent Rack 8 in the **Experiment** tab. An experiment is created automatically.
4. Tap to select sample position(s).
5. To assign a reagent to each sample, tap **Scan reagent** and scan the reagent barcode. Alternatively, tap on a free position of the MACS Reagent Rack 8 for selection out of the reagent list.
6. Unscrew the lids from the reagent vials and place the vials onto the designated positions on the MACS Reagent Rack 8.
7. Tap **Place reagent(s) on reagent rack** button in the dialog box.
8. Automated labeling is set automatically if autolabeling is supported and a reagent rack is selected. Alternatively, tap **Labeling** in the reagent placement dialog and select **Auto**.
9. Tap **Sample volume** in the **Sample process** pane and enter the sample volume. Tap the return key.
10. Tap **Run** to start the separation process.

2.6 (Optional) Removal of the REAlease Complex and second magnetic labeling with REAlease MicroBeads

▲ The recommended incubation temperature is at room temperature (19–25 °C).

▲ For second magnetic labeling with MACS Anti-Biotin MicroBeads proceed through all steps of chapter 2.5 to remove the REAlease Biotin Complex.

2.6.1 Removal of the REAlease Complex

1. Centrifuge cell suspension at 300×g for 10 minutes. Aspirate supernatant completely.
2. Resuspend cell pellet in appropriate amount of separation buffer:

   MS: 1 mL LS: 5 mL
3. Add an appropriate amount of REAlease Release Reagent:

   MS: 20 µL LS: 100 µL
4. Mix well and incubate for 5 minutes.
5. Cells are now free from REAlease Complex and MicroBeads and are ready for analysis or downstream applications.
6. (Optional) For second magnetic labeling with REAlease MicroBeads continue with 2.5.2.

2.6.2 Second magnetic labeling with REAlease MicroBeads

1. Centrifuge cell suspension at 300×g for 10 minutes. Aspirate supernatant completely.
2. Resuspend cell pellet in 40 µL of REAlease Stop Reagent per 10^7 total cells.
3. Mix well.
4. For a second magnetic labeling follow the labeling protocol in the respective REAlease MicroBead Kit data sheet.

   ▲ **Note:** For best recovery and purity of cells, the amount of MACS MicroBeads for the second positive labeling may need optimization as the starting frequency of target cells may be different from a PBMC sample.

2.7 (Optional) Second magnetic labeling with MACS MicroBeads

▲ For second magnetic labeling with MACS Anti-Biotin MicroBeads proceed through all steps of chapter 2.6 to remove the REAlease Biotin Complex.

1. Centrifuge cell suspension at 300×g for 10 minutes. Aspirate supernatant completely.

Unless otherwise specifically indicated, all Miltenyi Biotec products and services are for research use only and not for diagnostic or therapeutic use.

140-006-710.01


2. Add the recommended amount of MACS MicroBeads to label the cells magnetically for the second marker. For details refer to the respective MACS MicroBeads data sheet.

▲ **Note:** For best recovery and purity of cells, the amount of MACS MicroBeads for the second positive labeling may need optimization as the starting frequency of target cells may be different from a PBMC sample.

3. Example of a separation using the REAlease CD25 MicroBead Kit

CD25$^+$ cells were isolated from human PBMC using the REAlease CD25 MicroBead Kit, MS Columns, and a MiniMACS™ Separator. Cells were fluorescently stained with CD25-APC, CD4-VioBlue, and CD127-PE and analyzed by flow cytometry using the MACSQuant® Analyzer X. Cell debris and dead cells were excluded from the analysis based on scatter signals and propidium iodide fluorescence.

A) Cell purity




B) Bead-free cells: efficiency of REAlease Anti-Biotin MicroBeads release

Release efficiency was higher than 99% for the REAlease Anti-Biotin MicroBeads (CD25). The efficiency was determined by re-applying the isolated cells to a second MACS Column. The ratio between the numbers of cells in the flow-through and the total number of cells applied to the second column allowed us to calculate the efficiency of magnetic labeling removal.

C) Label-free cells: REAlease Biotin Complex release

The efficient removal of all labels was shown by using Biotin Antibody-FITC to analyze the cells by flow cytometry for the presence of REAlease Biotin Complex. Directly after isolation, the cells showed staining of biotin ("MicroBead-free CD25$^+$ cells"), whereas the label-free CD25$^+$ cells after the REAlease Biotin Complex release were negative for biotin similar to the non-labeled cells before separation.



Refer to **www.miltenyibiotec.com** for all data sheets and protocols. Miltenyi Biotec provides technical support worldwide. Visit www.miltenyibiotec.com for local Miltenyi Biotec Technical Support contact information.

Legal notices

Limited product warranty

Miltenyi Biotec B.V. & Co. KG and/or its affiliate(s) warrant this product to be free from material defects in workmanship and materials and to conform substantially with Miltenyi Biotec's published specifications for the product at the time of order, under normal use and conditions in accordance with its applicable documentation, for a period beginning on the date of delivery of the product by Miltenyi Biotec or its authorized distributor and ending on the expiration date of the product's applicable shelf life stated on the product label, packaging or documentation (as applicable) or, in the absence thereof, ONE (1) YEAR from date of delivery ("Product Warranty"). Miltenyi Biotec's Product Warranty is provided subject to the warranty terms as set forth in Miltenyi Biotec's General Terms and Conditions for the Sale of Products and Services available on Miltenyi Biotec's website at www.miltenyibiotec.com, as in effect at the time of order ("Product Warranty"). Additional terms may apply. BY USE OF THIS PRODUCT, THE CUSTOMER AGREES TO BE BOUND BY THESE TERMS.

THE CUSTOMER IS SOLELY RESPONSIBLE FOR DETERMINING IF A PRODUCT IS SUITABLE FOR CUSTOMER'S PARTICULAR PURPOSE AND APPLICATION METHODS.

Technical information

The technical information, data, protocols, and other statements provided by Miltenyi Biotec in this document are based on information, tests, or experience which Miltenyi Biotec believes to be reliable, but the accuracy or completeness of such information is not guaranteed. Such technical information and data are intended for persons with knowledge and technical skills sufficient to assess and apply their own informed judgment to the information. Miltenyi Biotec shall not be liable for any technical or editorial errors or omissions contained herein.

All information and specifications are subject to change without prior notice. Please contact Miltenyi Biotec Technical Support or visit www.miltenyibiotec.com for the most up-to-date information on Miltenyi Biotec products.

Licenses

This product and/or its use may be covered by one or more pending or issued patents and/or may have certain limitations. Certain uses may be excluded by separate terms and conditions. Please contact your local Miltenyi Biotec representative or visit Miltenyi Biotec's website at www.miltenyibiotec.com for more information.

The purchase of this product conveys to the customer the non-transferable right to use the purchased amount of the product in research conducted by the customer (whether the customer is an academic or for-profit entity). This product may not be further sold. Additional terms and conditions (including the terms of a Limited Use Label License) may apply.

CUSTOMER'S USE OF THIS PRODUCT MAY REQUIRE ADDITIONAL LICENSES DEPENDING ON THE SPECIFIC APPLICATION. THE CUSTOMER IS SOLELY RESPONSIBLE FOR DETERMINING FOR ITSELF WHETHER IT HAS ALL APPROPRIATE LICENSES IN PLACE. Miltenyi Biotec provides no warranty that customer's use of this product does not and will not infringe intellectual property rights owned by a third party. BY USE OF THIS PRODUCT, THE CUSTOMER AGREES TO BE BOUND BY THESE TERMS.

Trademarks

autoMACS, MACS, MACSQuant, MidiMACS, the Miltenyi Biotec logo, MiniMACS, OctoMACS, QuadroMACS, REAlease, SuperMACS, and Viobility are registered trademarks or trademarks of Miltenyi Biotec B.V. & Co. KG and/or its affiliates in various countries worldwide. All other trademarks mentioned in this publication are the property of their respective owners and are used for identification purposes only.

Ficoll-Paque is a trademark of GE Healthcare companies.

Copyright © 2023 Miltenyi Biotec and/or its affiliates. All rights reserved.